



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

Coherent, Inc.

27650 SW 95th Avenue

Wilsonville, OR 97070

(with satellite locations and capabilities identified on the scope of accreditation)

Fulfills the requirements of

ISO/IEC 17025:2017

and

ANSI/NCSL Z540-1-1994 (R2002)

In the field of

CALIBRATION

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

A handwritten signature in black ink, appearing to be 'Jason Stine', is positioned above a horizontal line.

Jason Stine, Vice President

Expiry Date: 15 May 2026

Certificate Number: AC-1630



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

**SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017
AND
ANSI/NCSL Z540-1-1994 (R2002)**

Coherent, Inc.
27650 SW 95th Avenue
Wilsonville, OR 97070
Philip Taylor 800-343-4912
Philip.Taylor@coherent.com

CALIBRATION

Valid to: **May 15, 2026**

Certificate Number: **AC-1630**

Photometry and Radiometry

| Parameter / Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method and/or Equipment |
|--|--|--|---|
| Electrical Calibration of Laser Power Thermal Meters | 100 μ W to 5 000 W 1.831 μ V to 2 V | 0.8 % of reading | ARB, DMM, Attenuator |
| Electrical Calibration of Laser Power Optical Meters | 10 nW to 30 mW 51 nA to 19.1 mA | 0.8 % of reading | Current Source |
| Electrical Calibration of Laser Energy Meters | 100 nJ to 3 J 40 μ V to 5 V | 0.8 % of reading | ARB, Attenuator |
| Laser Power Measuring Sensors at 514 nm | (0.18 to 1.2) W (0.000 2 to 7) V/W | 1 % of reading 1 % of reading | Coherent Working Standard Sensor, DMM, Coherent Power Meter |
| Laser Power Measuring kW Sensors at 1070 (+/-10) nm | (360 to 440) W (0.000 004 5 to 0.000 4) V/W | 2.5 % of reading 2.5 % of reading | Coherent Working Standard Sensor, DMM, Coherent Power Meter |
| Laser Power Measuring kW Sensors at 10 600 nm | 200 W to 1.1 kW (0.000 06 to 0.000 4) V/W | 3.7 % of reading 3.7 % of reading | Coherent Working Standard Sensor, DMM, Coherent Power Meter |
| Energy Measuring Sensors at 193 nm | 3 μ J to 2 mJ (24 to 80) V/J | 2 % of reading 2 % of reading | Coherent Working Standard Sensor, Oscilloscope, Coherent Energy Meter |
| Energy Measuring Sensors at 248 nm | 60 μ J to 7 mJ (6 to 32) V/J | 2.1 % of reading 2.1 % of reading | |
| Energy Measuring Sensors at 1 064 nm | 10 μ J to 160 mJ (2 to 21 700) V/J | 2 % of reading 2 % of reading | |



ANSI National Accreditation Board

Photometry and Radiometry

| Parameter / Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method and/or Equipment |
|--|---|--|---|
| Laser Power Measuring Sensors at 810 (+/- 10) nm | (70 to 80) W (0.08 to 13) mV/W | 1.7 % of reading | Coherent Working Standard Sensor, Coherent Power Meter |
| Laser Power Measuring Sensors at 10 600 nm | (0.5 to 150) W (0.001 4 to 0.226) V/W | 2 % of reading 2 % of reading | Coherent Working Standard Sensor, Coherent Power Meter |
| Laser Power Measuring kW Sensors at 1070 (+/-10) nm | (250 to 12 000) W (0.000 004 5 to 0.000 4) V/W | 2.5 % of reading 2.5 % of reading | |
| Laser Power Measuring Sensors from 250 nm to 400 nm (UV-Range) | 200 nW to 20 μ W (0.05 to 0.2) A/W | 3 % of reading 3 % of reading | Coherent Working Standard Sensor, DMM, Coherent Power Meter |
| Laser Power Measuring Sensors from 400 nm to 450 nm (VIS-Range) | 3 μ W to 20 μ W (0.01 to 0.1) A/W | 10 % of reading 10 % of reading | |
| Laser Power Measuring Sensors from 450 nm to 1100 nm (VIS-Range) | 3 μ W to 100 μ W (0.01 to 1) A/W | 4.8 % of reading 4.8 % of reading | |
| Laser Power Measuring Sensors from 800 nm to 1 700 nm (IR-Range) | 300 nW to 60 μ W (0.1 to 1.2) A/W | 4 % of reading 4 % of reading | |
| Laser Power Measuring Sensors from 1 700 nm to 1 800 nm (IR-Range) | 200 nW to 1 μ W (0.08 to 1) A/W | 8 % of reading 8 % of reading | |



ANSI National Accreditation Board

**Services performed at satellite laboratory:
Coherent (Deutschland) GmbH**

Dieselstrasse 5B
Dieburg, Germany D-64807
Philip Taylor 800-343-4912
Philip.Taylor@coherent.com

Photometry and Radiometry

| Parameter / Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method and/or Equipment |
|--|--|---|---|
| Electrical Calibration of Laser Power Thermal Meters | 100 μ W to 5 000 W 1.831 μ V to 2 V | 0.8 % of reading | ARB, DMM, Attenuator |
| Electrical Calibration of Laser Power Optical Meters | 10 nW to 30 mW 51 nA to 19.1 mA | 0.8 % of reading | Current Source |
| Electrical Calibration of Laser Energy Meters | 100 nJ to 3 J 40 μ V to 5 V | 0.8 % of reading | ARB, Attenuator |
| Laser Power Measuring Sensors at 514 nm | (0.18 to 1.2) W (0.000 2 to 7) V/W | 1 % of reading 1 % of reading | Coherent Working Standard Sensor, DMM, Coherent Power Meter |
| Laser Power Measuring Sensors at 10 600 nm | (0.5 to 150) W (0.001 4 to 0.226) V/W | 2 % of reading 2 % of reading | Coherent Working Standard Sensor, Coherent Power Meter |
| Energy Measuring Sensors at 193 nm | 3 μ J to 2 mJ (24 to 80) V/J | 2 % of reading 2 % of reading | Coherent Working Standard Sensor, Oscilloscope, Coherent Energy Meter |
| Energy Measuring Sensors at 1 064 nm | 10 μ J to 160 mJ (2 to 21 700) V/J | 2 % of reading 2 % of reading | |
| Laser Power Measuring Sensors from 250 nm to 400 nm (UV- Range) | 200 nW to 20 μ W (0.05 to 0.2) A/W | 3 % of reading 3 % of reading | Coherent Working Standard Sensor, DMM, Coherent Power Meter |
| Laser Power Measuring Sensors from 400 nm to 450 nm (VIS- Range) | 3 μ W to 20 μ W (0.01 to 0.1) A/W | 10 % of reading 10 % of reading | |
| Laser Power Measuring Sensors from 450 nm to 1 100 nm (VIS- Range) | 3 μ W to 100 μ W (0.01 to 1) A/W | 4.8 % of reading 4.8 % of reading | |
| Laser Power Measuring Sensors from 800 nm to 1 700 nm (IR- Range) | 300 nW to 60 μ W (0.1 to 1.2) A/W | 4 % of reading 4 % of reading | Coherent Working Standard Sensor, DMM, Coherent Power Meter |

Photometry and Radiometry

| Parameter / Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method and/or Equipment |
|---|--|---|---|
| Laser Power Measuring Sensors from 1 700 nm to 1 800 nm (IR- Range) | 200 nW to 1 μ W (0.08 to 1) A/W | 8 % of reading 8 % of reading | Coherent Working Standard Sensor, DMM, Coherent Power Meter |

Services performed at satellite laboratory:

Coherent Japan, Inc.

Business Office: 26F Shinjuku Maynds Tower
2-1-1 Yoyogi, Shibuya-ku Tokyo, Japan 151-0053
Service Location: Atsugi Tech Center
1042-4 Toda, Atsugi-shi, Kanagawa, Japan 243-0023
Philip.Taylor@coherent.com

Photometry and Radiometry

| Parameter / Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method and/or Equipment |
|---|--|---|---|
| Electrical Calibration of Laser Power Thermal Meters | 100 μ W to 5 000 W 1.831 μ V to 2 V | 0.8 % of reading | ARB, DMM, Attenuator |
| Electrical Calibration of Laser Power Optical Meters | 10 nW to 30 mW 51 nA to 19.1 mA | 0.8 % of reading | Current Source |
| Electrical Calibration of Laser Energy Meters | 100 nJ to 3 J 40 μ V to 5 V | 0.8 % of reading | ARB, Attenuator |
| Laser Power Measuring Sensors at 514 nm | (0.18 to 1.2) W (0.000 2 to 7) V/W | 1 % of reading 1 % of reading | Coherent Working Standard Sensor, DMM, Coherent Power Meter |
| Laser Power Measuring Sensors at 10 600 nm | (0.5 to 150) W (0.001 4 to 0.226) V/W | 2 % of reading 2 % of reading | Coherent Working Standard Sensor, Coherent Power Meter |
| Laser Power Measuring Sensors from 250 nm to 400 nm (UV-Range) | 200 nW to 20 μ W (0.05 to 0.2) A/W | 3 % of reading 3 % of reading | Coherent Working Standard Sensor, DMM, Coherent Power Meter |
| Laser Power Measuring Sensors from 400 nm to 450 nm (VIS-Range) | 3 μ W to 20 μ W (0.01 to 0.1) A/W | 10 % of reading 10 % of reading | Coherent Working Standard Sensor, DMM, Coherent Power Meter |


Photometry and Radiometry

| Parameter / Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method and/or Equipment |
|--|---|---|---|
| Laser Power Measuring Sensors from 450 nm to 1 100 nm (VIS-Range) | 3 μ W to 100 μ W (0.01 to 1) A/W | 4.8 % of reading 4.8 % of reading | Coherent Working Standard Sensor, DMM, Coherent Power Meter |
| Laser Power Measuring Sensors from 800 nm to 1 700 nm (IR-Range) | 300 nW to 60 μ W (0.1 to 1.2) A/W | 4 % of reading 4 % of reading | Coherent Working Standard Sensor, DMM, Coherent Power Meter |
| Laser Power Measuring Sensors from 1 700 nm to 1 800 nm (IR-Range) | 200 nW to 1 μ W (0.08 to 1) A/W | 8 % of reading 8 % of reading | Coherent Working Standard Sensor, DMM, Coherent Power Meter |

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-1630.



Jason Stine, Vice President